

PERFUSION ADAPTER SET

R_x Only

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Intended Use/Indications:

The Perfusion Adapter Set is indicated for use when 3/8" ID or 1/4" ID tubing is to be adapted for intravenous infusion, for example, following cardiopulmonary bypass.

Contraindications:

There are no known contraindications specific to this device.



Precautions:

Caution: Always ensure the connection between the tubing and Perfusion Adapter Set is secure to avoid blood loss or possible air embolism.

Caution: Always remove air from the Perfusion Adapter Set prior to use to avoid air embolism.

Caution: Inspect sealed sterile packaging before opening. If seal is broken, contents may not be sterile and may cause infection in the patient.

Caution: Sets with "Y" medication injection sites are designed for use with 21 gauge or smaller needles to avoid damaging the seal.

Caution: For single patient use only. Do not reuse, reprocess or re-sterilize. Reuse, reprocessing or re-sterilization will compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury or illness.

Direction for Use:

1. Attach Perfusion Adapter Set to any suitable infusion cannula or catheter and secure all connections.
2. Prime and de-bubble the Perfusion Adapter Set prior to beginning infusion.



NovoSci[™]

2021 Airport Rd.
Conroe, TX 77301
Toll Free 1-800-854-0567
Direct 1-281-363-4949

424007-000 Rev. L



Contains or presence of Phthalates:
Bis (2-ethylhexyl) phthalate (DEHP). The potential effects of phthalates on pregnant/ nursing women or children have not been fully characterized and there may be concern for reproductive and development effects.

PERFUSION ADAPTER SET

R_x Only

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Intended Use/Indications:

The Perfusion Adapter Set is indicated for use when 3/8" ID or 1/4" ID tubing is to be adapted for intravenous infusion, for example, following cardiopulmonary bypass.

Contraindications:

There are no known contraindications specific to this device.



Precautions:

Caution: Always ensure the connection between the tubing and Perfusion Adapter Set is secure to avoid blood loss or possible air embolism.

Caution: Always remove air from the Perfusion Adapter Set prior to use to avoid air embolism.

Caution: Inspect sealed sterile packaging before opening. If seal is broken, contents may not be sterile and may cause infection in the patient.

Caution: Sets with "Y" medication injection sites are designed for use with 21 gauge or smaller needles to avoid damaging the seal.

Caution: For single patient use only. Do not reuse, reprocess or re-sterilize. Reuse, reprocessing or re-sterilization will compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury or illness.

Direction for Use:

1. Attach Perfusion Adapter Set to any suitable infusion cannula or catheter and secure all connections.
2. Prime and de-bubble the Perfusion Adapter Set prior to beginning infusion.



NovoSci[™]

2021 Airport Rd.
Conroe, TX 77301
Toll Free 1-800-854-0567
Direct 1-281-363-4949

424007-000 Rev. L



Contains or presence of Phthalates:
Bis (2-ethylhexyl) phthalate (DEHP). The potential effects of phthalates on pregnant/ nursing women or children have not been fully characterized and there may be concern for reproductive and development effects.