

QUICKIE PRIME PERFUSATE INFUSION SET

R_x Only

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Intended Use/Indications:

The Quickie Prime Perfusate Infusion Set is used to rapidly prime an extracorporeal circuit through an indirect 1/4" inlet connector of reservoir.

Contraindications:

There are no known contraindications specific to this device.



Warnings and Precautions:

Warning: Do not connect Quickie Prime Perfusate Infusion Set to an inlet connector that empties directly into the open chamber of the reservoir during a cardiopulmonary bypass as air entrainment may result in air embolism.

Caution: Inspect sealed sterile packaging before opening. If seal is broken, contents may not be sterile and may cause infection in the patient.

Caution: Sets with "Y" medication injection sites are designed for use with 21 gauge or smaller needles to avoid damaging the seal.

Caution: For single patient use only. Do not reuse, reprocess or re-sterilize. Reuse, reprocessing or re-sterilization will compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury or illness.

Direction for Use:

1. Connect Quickie Prime to appropriate filtered inlet.
2. Remove the I.V. spike protector from the Quickie Prime and insert into the appropriate outlet port of the perfusate container.



NovoSci™

2021 Airport Rd.
Conroe, TX 77301
Toll Free 1-800-854-0567
Direct 1-281-363-4949

424008-000 Rev. M



Contains or presence of Phthalates:
Bis (2-ethylhexyl) phthalate (DEHP). The potential effects of phthalates on pregnant/ nursing women or children have not been fully characterized and there may be concern for reproductive and development effects.

QUICKIE PRIME PERFUSATE INFUSION SET

R_x Only

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Intended Use/Indications:

The Quickie Prime Perfusate Infusion Set is used to rapidly prime an extracorporeal circuit through an indirect 1/4" inlet connector of reservoir.

Contraindications:

There are no known contraindications specific to this device.



Warnings and Precautions:

Warning: Do not connect Quickie Prime Perfusate Infusion Set to an inlet connector that empties directly into the open chamber of the reservoir during a cardiopulmonary bypass as air entrainment may result in air embolism.

Caution: Inspect sealed sterile packaging before opening. If seal is broken, contents may not be sterile and may cause infection in the patient.

Caution: Sets with "Y" medication injection sites are designed for use with 21 gauge or smaller needles to avoid damaging the seal.

Caution: For single patient use only. Do not reuse, reprocess or re-sterilize. Reuse, reprocessing or re-sterilization will compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury or illness.

Direction for Use:

1. Connect Quickie Prime to appropriate filtered inlet.
2. Remove the I.V. spike protector from the Quickie Prime and insert into the appropriate outlet port of the perfusate container.



NovoSci™

2021 Airport Rd.
Conroe, TX 77301
Toll Free 1-800-854-0567
Direct 1-281-363-4949

424008-000 Rev. M



Contains or presence of Phthalates:
Bis (2-ethylhexyl) phthalate (DEHP). The potential effects of phthalates on pregnant/ nursing women or children have not been fully characterized and there may be concern for reproductive and development effects.