

VESSEL CLAMP

R_x Only

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Intended Use/Indications:

The Vessel Clamp is designed to be used for transient minimally traumatic occlusion of blood vessels in coronary artery graft surgery where gentle clamping pressure is desired to avoid excessive trauma to the vessel.

Contraindications:

There are no known contraindications specific to this device.



Precautions:

Caution: Inspect sealed sterile packaging before opening. If seal is broken, contents may not be sterile and may cause infection in the patient.

Caution: For single patient use only. Do not reuse, reprocess or re-sterilize. Reuse, reprocessing or re-sterilization will compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury or illness.

Direction for Use:

1. Remove Vessel Clamp from package in its “open” position.
2. Realign jaws so that Vessel Clamp is in its “closed” position prior to use.



NovoSci[™]

2021 Airport Rd.
Conroe, TX 77301
Toll Free 1-800-854-0567
Direct 1-281-363-4949

424010-000 Rev. L



Contains or presence of Phthalates:
Bis (2-ethylhexyl) phthalate (DEHP). The potential effects of phthalates on pregnant/ nursing women or children have not been fully characterized and there may be concern for reproductive and development effects.

VESSEL CLAMP

R_x Only

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Intended Use/Indications:

The Vessel Clamp is designed to be used for transient minimally traumatic occlusion of blood vessels in coronary artery graft surgery where gentle clamping pressure is desired to avoid excessive trauma to the vessel.

Contraindications:

There are no known contraindications specific to this device.



Precautions:

Caution: Inspect sealed sterile packaging before opening. If seal is broken, contents may not be sterile and may cause infection in the patient.

Caution: For single patient use only. Do not reuse, reprocess or re-sterilize. Reuse, reprocessing or re-sterilization will compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury or illness.

Direction for Use:

1. Remove Vessel Clamp from package in its “open” position.
2. Realign jaws so that Vessel Clamp is in its “closed” position prior to use.



NovoSci[™]

2021 Airport Rd.
Conroe, TX 77301
Toll Free 1-800-854-0567
Direct 1-281-363-4949

424010-000 Rev. L



Contains or presence of Phthalates:
Bis (2-ethylhexyl) phthalate (DEHP). The potential effects of phthalates on pregnant/ nursing women or children have not been fully characterized and there may be concern for reproductive and development effects.