

DISPOSABLE ARTERIAL FILTER PURGE LINE

R_x Only

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Intended Use/Indications:

The Purge Line is indicated for us to provide continuous purging of air from an arterial line filter during cardiopulmonary bypass.

Contraindications:

The Purge Line is not, in itself, a medical treatment. There are, therefore, no absolute contraindications other than those for the cardiopulmonary bypass in general.



Precautions:

Caution: Leakage at connections at any point in this line can cause loss of sterility, fluid and/or blood loss, or air embolism.

Caution: The Purge Line is disposable. Do not reuse, reprocess or re-sterilize.

Direction for Use:

1. Aseptically remove the Purge Line from the peel pouch and transfer it to a sterile field.
2. Remove end cap from male locking luer on the stopcock and attach securely to the female port on the arterial filter, using a clockwise push/twist motion.
3. Remove end cap from male locking luer on the opposite end of line and attach securely, using a clockwise push/twist motion, to an appropriate component in the extracorporeal circuit.
4. Visually inspect purge line to ensure that its free of kinks, air bubbles and leaks before initiating cardiopulmonary bypass.

DISPOSABLE ARTERIAL FILTER PURGE LINE

R_x Only

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Intended Use/Indications:

The Purge Line is indicated for us to provide continuous purging of air from an arterial line filter during cardiopulmonary bypass.

Contraindications:

The Purge Line is not, in itself, a medical treatment. There are, therefore, no absolute contraindications other than those for the cardiopulmonary bypass in general.



Precautions:

Caution: Leakage at connections at any point in this line can cause loss of sterility, fluid and/or blood loss, or air embolism.

Caution: The Purge Line is disposable. Do not reuse, reprocess or re-sterilize.

Direction for Use:

1. Aseptically remove the Purge Line from the peel pouch and transfer it to a sterile field.
2. Remove end cap from male locking luer on the stopcock and attach securely to the female port on the arterial filter, using a clockwise push/twist motion.
3. Remove end cap from male locking luer on the opposite end of line and attach securely, using a clockwise push/twist motion, to an appropriate component in the extracorporeal circuit.
4. Visually inspect purge line to ensure that its free of kinks, air bubbles and leaks before initiating cardiopulmonary bypass.



NovoSci[™]

2021 Airport Rd.
Conroe, TX 77301
Toll Free 1-800-854-0567
Direct 1-281-363-4949

424644-000 Rev. H



NovoSci[™]

2021 Airport Rd.
Conroe, TX 77301
Toll Free 1-800-854-0567
Direct 1-281-363-4949

424644-000 Rev. H