

NovoSci Extracorporeal Tube Pack

Directions for Use:

1. Remove product from package.
2. Connect to the extracorporeal circuit observing sterile technique.
3. Prime the system in accordance with the manufacturer's instructions for all components in the extracorporeal circuit.
4. Recirculation and microfiltration of the priming solution is recommended.
5. Inspect the extracorporeal circuit for air bubbles or leaks before initiating cardiopulmonary bypass.
6. Initiate cardiopulmonary bypass in accordance with the manufacturer's instructions of all components in the extracorporeal circuit.

Cautions and Warnings:

7. The user is responsible for any changes in the cardiopulmonary bypass circuit or procedure that would compromise its function.
8. Leakage at any point in the extracorporeal circuit can cause blood loss or air embolism. Careful observation of the circuit for leaks before and during cardiopulmonary bypass is recommended.
9. Do not re-sterilize the pack or any component in it.
10. Do not reuse any part of the product. It is disposable and intended for one use only.
11. Improper occlusion of roller pumps can cause pump head failure. Pump head occlusion should be accomplished in accordance with the pump manufacturer's recommended procedure.
12. If silicone rubber pump segments are used, do not lubricate with oils or greases. Tubing failures may result.

CAUTION: FEDERAL LAW RESTRICTS THIS DEVICE TO SALE BY OR ON THE ORDER OF A PHYSICIAN.

Manufactured and distributed by:



NovoSci
2021 Airport Road
Conroe, Texas 77301
Toll Free 1-800-854-0567
In TX (281) 363-4949



Contains or Presence of Phthalates: Bis (2-ethylhexyl) phthalate (DEHP). The potential effects of phthalates on

Pregnant/nursing women or children have not been fully characterized and there may be concern for reproductive and development effects.

R-492 Rev. O

NovoSci Extracorporeal Tube Pack

Directions for Use:

1. Remove product from package.
2. Connect to the extracorporeal circuit observing sterile technique.
3. Prime the system in accordance with the manufacturer's instructions for all components in the extracorporeal circuit.
4. Recirculation and microfiltration of the priming solution is recommended.
5. Inspect the extracorporeal circuit for air bubbles or leaks before initiating cardiopulmonary bypass.
6. Initiate cardiopulmonary bypass in accordance with the manufacturer's instructions of all components in the extracorporeal circuit.

Cautions and Warnings:

7. The user is responsible for any changes in the cardiopulmonary bypass circuit or procedure that would compromise its function.
8. Leakage at any point in the extracorporeal circuit can cause blood loss or air embolism. Careful observation of the circuit for leaks before and during cardiopulmonary bypass is recommended.
9. Do not re-sterilize the pack or any component in it.
10. Do not reuse any part of the product. It is disposable and intended for one use only.
11. Improper occlusion of roller pumps can cause pump head failure. Pump head occlusion should be accomplished in accordance with the pump manufacturer's recommended procedure.
12. If silicone rubber pump segments are used, do not lubricate with oils or greases. Tubing failures may result.

CAUTION: FEDERAL LAW RESTRICTS THIS DEVICE TO SALE BY OR ON THE ORDER OF A PHYSICIAN.

Manufactured and distributed by:



NovoSci
2021 Airport Road
Conroe, Texas 77301
Toll Free 1-800-854-0567
In TX (281) 363-4949



Contains or Presence of Phthalates: Bis (2-ethylhexyl) phthalate (DEHP). The potential effects of phthalates on

Pregnant/nursing women or children have not been fully characterized and there may be concern for reproductive and development effects.

R-492 Rev. O

NovoSci Extracorporeal Tube Pack

Directions for Use:

1. Remove product from package.
2. Connect to the extracorporeal circuit observing sterile technique.
3. Prime the system in accordance with the manufacturer's instructions for all components in the extracorporeal circuit.
4. Recirculation and microfiltration of the priming solution is recommended.
5. Inspect the extracorporeal circuit for air bubbles or leaks before initiating cardiopulmonary bypass.
6. Initiate cardiopulmonary bypass in accordance with the manufacturer's instructions of all components in the extracorporeal circuit.

Cautions and Warnings:

7. The user is responsible for any changes in the cardiopulmonary bypass circuit or procedure that would compromise its function.
8. Leakage at any point in the extracorporeal circuit can cause blood loss or air embolism. Careful observation of the circuit for leaks before and during cardiopulmonary bypass is recommended.
9. Do not re-sterilize the pack or any component in it.
10. Do not reuse any part of the product. It is disposable and intended for one use only.
11. Improper occlusion of roller pumps can cause pump head failure. Pump head occlusion should be accomplished in accordance with the pump manufacturer's recommended procedure.
12. If silicone rubber pump segments are used, do not lubricate with oils or greases. Tubing failures may result.

CAUTION: FEDERAL LAW RESTRICTS THIS DEVICE TO SALE BY OR ON THE ORDER OF A PHYSICIAN.

Manufactured and distributed by:



NovoSci
2021 Airport Road
Conroe, Texas 77301
Toll Free 1-800-854-0567
In TX (281) 363-4949



Contains or Presence of Phthalates: Bis (2-ethylhexyl) phthalate (DEHP). The potential effects of phthalates on

Pregnant/nursing women or children have not been fully characterized and there may be concern for reproductive and development effects.

R-492 Rev. O

NovoSci Extracorporeal Tube Pack

Directions for Use:

1. Remove product from package.
2. Connect to the extracorporeal circuit observing sterile technique.
3. Prime the system in accordance with the manufacturer's instructions for all components in the extracorporeal circuit.
4. Recirculation and microfiltration of the priming solution is recommended.
5. Inspect the extracorporeal circuit for air bubbles or leaks before initiating cardiopulmonary bypass.
6. Initiate cardiopulmonary bypass in accordance with the manufacturer's instructions of all components in the extracorporeal circuit.

Cautions and Warnings:

7. The user is responsible for any changes in the cardiopulmonary bypass circuit or procedure that would compromise its function.
8. Leakage at any point in the extracorporeal circuit can cause blood loss or air embolism. Careful observation of the circuit for leaks before and during cardiopulmonary bypass is recommended.
9. Do not re-sterilize the pack or any component in it.
10. Do not reuse any part of the product. It is disposable and intended for one use only.
11. Improper occlusion of roller pumps can cause pump head failure. Pump head occlusion should be accomplished in accordance with the pump manufacturer's recommended procedure.
12. If silicone rubber pump segments are used, do not lubricate with oils or greases. Tubing failures may result.

CAUTION: FEDERAL LAW RESTRICTS THIS DEVICE TO SALE BY OR ON THE ORDER OF A PHYSICIAN.

Manufactured and distributed by:



NovoSci
2021 Airport Road
Conroe, Texas 77301
Toll Free 1-800-854-0567
In TX (281) 363-4949



Contains or Presence of Phthalates: Bis (2-ethylhexyl) phthalate (DEHP). The potential effects of phthalates on

Pregnant/nursing women or children have not been fully characterized and there may be concern for reproductive and development effects.

R-492 Rev. O